



Lab. N.1962L
Strada del Petriccio e Belriguardo 35, Siena 53100
P.iva 01504290527 | www.ambra.life

Certificate of Analysis # 2024-2608/EXT/HPLC

Informations provided by the client

Sample Name: CHARAS CBD
Matrix: Plant Material
Product: Resin

Laboratory Information

Acceptance Date: 14/10/2024
Test Start Date: 15/10/2024
Sampling: Sample delivered by the customer
Sample ID: H1410242496/EXT

Prepared for

CBWeed srl
viale Roma 203/A Forlì
VAT 04344770401

Compound	Result	UM	LOQ	Result (mg/g)	Measurement uncertainty
CBDV	< LOQ	%(m/m)	0,02	< LOQ	
CBDA	1,03	%(m/m)	0,033	10,3	
CBGA	7,76	%(m/m)	0,035	77,6	
CBG	5,55	%(m/m)	0,018	55,5	
CBD	18,75	%(m/m)	0,015	187,5	
THCV	< LOQ	%(m/m)	0,02	< LOQ	
CBN	14,53	%(m/m)	0,02	145,3	
d9-THC	0,07	%(m/m)	0,028	0,7	
CBC	0,19	%(m/m)	0,017	1,9	
THCA	< LOQ	%(m/m)	0,02	< LOQ	
TOTAL CBD	19,65	%(m/m)	0,02	196,5	
TOTAL CBG	12,36	%(m/m)	0,02	123,6	
TOTAL THC	0,07	%(m/m)	0,02	0,7	
TOTAL MOISTURE	4,35	%	0,05	43,5	

%(m/m) = (Mass of the analyte/Mass of the product as it is)
NR Not Detected
LOD Limit of detection
LOQ Limit of quantification
<LOQ Below the limit of quantification

Reason for Review:
Revisione risultati

This revision cancels and replaces the previously issued document.

IST01 REV03 2022 (HPLC)
IST16 REV02 2022 (Thermogravimetry)

Total Cannabinoids are calculated using the following formulas to calculate the loss of the carboxyl group during decarboxylation
TOTAL THC = THC + (THCA * 0.877)
TOTAL CBD = CBD + (CBDA * 0.877)
TOTAL CBG = CBG + (CBGA * 0.877)

2496/EXT



End of Test Report # Certificate of Analysis # 2024-2608/EXT/HPLC

Test end date: 16/10/2024
Issuing date: 16/10/2024

Firmato da:
DAVIDE DE ROSSI
Responsabile del Laboratorio

The results reported in the Test Report refer only to the sample actually analyzed. The Laboratory declines all responsibility for the methods of carrying out sampling, transport and storage of the samples until delivery, if carried out by the customer. The Laboratory assumes responsibility for the information and data contained in the test report, with the exclusion of what is declared by the customer. This CoA can only be reproduced in complete form: if not complete, it is allowed only after written authorization from the Laboratory Manager. Copies of this Test Report and related documents are kept for 4 years. The laboratory declines responsibility for any information provided by the customer that may affect the validity of the results. Test Report digitally signed and compliant pursuant to art. 23 Legislative Decree 7 March 2005 No. 82 CAD and subsequent amendments and additions.